

VISU Top

Light, adjustable and protective

VISU Top safety cabinet can be placed on a table.

- Owing to the transparent sides and front glass, no separate light is required.
- The sash can be adjusted to the desired height, with the suction working efficiently and any splashes remain inside. The sash is easy to operate owing to the counterweight mechanism.
- The fume hood is always equipped as requested by the customer. Our experts will be happy to help you choose the right solutions.




Product information

| | |
|----------------------------|---|
| Standard widths | 900, 1200 and 1500 mm |
| Work area depth | 550, 650 and 750 mm |
| Total depth | 630, 730 and 830 mm |
| Total height | Max. 1780 mm sash opened. Varies by width, more details in the Additional technical information tab. |
| Side materials | Clear polycarbonate. |
| Front wall | Vertical or diagonal laminated glass. |
| Back wall materials | High-pressure laminate. |
| Frame | Powder coated steel / high-pressure laminate-coated, humidity-resistant chipboard, also available in metal only. |
| By order | Stand rods, 4 pcs, made of acidproof steel; perforated plate shelves made of acidproof steel, also available with height adjustment, drying rack. |

VISU KALUSTE OY

 Eeliksentie 6, 34870 Visuvesi, Finland

 visu@visu.fi

 +358 44 749 4940

 visu.fi



INNOSTAVIA YMPÄRISTÖJÄ
TIETEELLE JA TUTKIMUKSELLE

Additional technical information


The technical properties depend on the hood width.

| Standard widths [mm] | 900 | 1200 | 1500 | |
|---------------------------------------|------|------|------|------------------------------------|
| Interior width [mm] | 820 | 1120 | 1420 | Depending on side materials. |
| Total height [mm] | 1780 | 1700 | 1600 | Sash opened |
| Air volume [l/s] | 103 | 140 | 178 | The sash 250mm open, 0.5 m/s flow. |
| Air conditioning pipe connection [mm] | 160 | 200 | 250 | |


Delivery and installation

Safety cabinets are delivered, if so agreed, installed in place, ready to be connected to building technology systems. A safety cabinet must be connected to a separate exhaust air duct. Any water supply, drainage, gas and electricity work must be performed by professionals.

VISU KALUSTE OY

 Eeliksentie 6, 34870 Visuvesi, Finland

 visu@visu.fi

 +358 44 749 4940

 visu.fi